

Agilent Cary FTIR Oil Analyzer

User's Guide



Agilent Technologies

Notices

© Agilent Technologies, Inc. 2008, 2010 and 2011

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Agilent Technologies, Inc. as governed by United States and international copyright laws.

Manual Part Number

8510248600

Edition

Fourth edition, November 2011

Printed in Australia

Agilent Technologies, Inc.

Warranty

The material contained in this document is provided “as is,” and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Agilent disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Agilent shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Agilent and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license.

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or subcontract, Software is delivered and licensed as “Commercial computer software” as defined in DFAR 252.227-7014 (June 1995), or as a “commercial item” as defined in FAR 2.101(a) or as “Restricted computer software” as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or

contract clause. Use, duplication or disclosure of Software is subject to Agilent Technologies’ standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

Contents


1. Safety Practices and Hazards	7
General	8
Electrical hazards	8
Laser safety	10
Labeling	11
Other precautions	12
Warning and caution messages	13
Information symbols	14
Color coding	15
CE compliance	16
Electromagnetic Compatibility	16
EN55011/CISPR11	16
ICES/NMB-001	17
2. Introduction	19
Specifications	21
Power	22
Laser	23
Weights and dimensions	24
Installation requirements	24
Laboratory	25
Computer	26
Training	26

Contents

Documentation	27
Conventions	28
3. Unpacking and Installation	29
Unpacking	29
Installation space requirement	31
Setting up	31
Spectrometer	31
Pump	33
Sample probe	36
Zinc selenide sample cell	36
Sampling plumbing	38
Installing Resolutions Pro	40
Software	40
Help	42
4. Operation	43
Power connections	44
Spectrometer	44
Peristaltic pump(s)	45
Spectrometer warm-up	46
Loading samples	47
Cleaning the cell	48
Starting the software	49
Turning off the system	50

5. Advanced Autosampler	51
Installation space requirement	52
Setting up	53
Checking	56
Bottle numbering convention	58
6. Maintenance	59
Routine	60
Cleaning	61
Spare parts	61
7. Troubleshooting	63

This page is intentionally left blank.



1. Safety Practices and Hazards

General	8
Electrical hazards	8
Laser safety	10
Other precautions	12
Warning and caution messages	13
Information symbols	14
Color coding	15
CE compliance	16
Electromagnetic Compatibility	16

Your Agilent Cary 600 Series FTIR Oil Analyzer has been carefully designed so that when used properly you have an accurate, fast, flexible and safe system.

If the equipment is used in a manner not specified by the manufacturer, the protection provided by the equipment may be impaired.

Information about safety practices appears throughout the documentation (both hard copy and online) provided with your instrument and accessories. Before using the Agilent Oil Analysis accessory, you must thoroughly read these safety practices and the 'Safety Practices and Hazards' section in the manual supplied with your Agilent FTIR spectrometer.

Observe all relevant safety practices at all times.

General

Operation of an Agilent Cary FTIR spectrometer fitted with an Oil Analysis accessory involves the use of high voltage energy, hazardous materials, and a helium-neon Class 2 laser operating in the visible region at 632.8 nanometers. Careless, improper, or unskilled use of your spectrometer or its accessories can cause death or serious injury to personnel, and/or severe damage to equipment and property.

The spectrometer incorporates interlocks and covers that are designed to prevent inadvertent contact with any potential hazards. If the instrument is used in any manner not specified by Agilent, this protection may be impaired. Develop safe working habits that do not depend upon the correct operation of the interlocks for safe operation. Do not bypass any interlock or cover.

These safety practices are provided to help you operate the instrument safely. Read each safety topic thoroughly before attempting to operate the instrument and **always** operate the spectrometer and its accessories in accordance with these safety practices.

Electrical hazards

The Agilent Cary FTIR spectrometers, the Oil Analysis accessory and other ancillary equipment contain electrical circuits, devices, and components operating at dangerous voltages. Contact with these circuits, devices, and components can cause death, serious injury, or painful electrical shock.

Good grounding/earthing is essential to avoid a potentially serious electric shock hazard. Ensure that there is an integral ground connection between the metal base of the instrument and accessories and the 3-pin earth-grounded mains-socket outlet. Consult the manual(s) and/or product labels supplied with your computer, monitor and printer for the relevant grounding requirements.

NOTE

This model is Equipment Class I. Application of the wrong supply voltage can create a fire hazard and a potentially serious shock hazard, and could seriously damage the FTIR spectrometer, its accessories and any attached ancillary equipment.

NOTE

The Cary 600 Series FTIR spectrometers, and the optional Advanced Autosampler, have a universal power supply that adapts to the supply voltage. However, care must be taken to ensure that a suitable AC (alternating current) voltage is used.

The peristaltic pump that is supplied with the Oil Analysis accessory, or the optional Advanced Autosampler (see Chapter 5), is designed for 100-240 VAC; 47-63 Hz operation.

CAUTION

If an FTIR Oil Analyzer is to be relocated to another site that has a different VAC supply, Agilent recommends that you contact your local representative to ensure you have the correct power cables and ensure you meet any requirement for a grounded converter or transformer for the pump(s) on the system.

Consult the manual(s) supplied with your computer, monitor and printer for their specific voltage requirements.

Replace blown fuses with fuses of the size and rating as stipulated in the text adjacent to the fuse holder or refer to the specification cited in the corresponding manual.

Do not use power cords with faulty or frayed insulation.

CAUTION

Use of controls or adjustments or performance or procedures other than those specified in this manual, may result in hazardous radiation exposure.

Laser safety

The Agilent Cary 600 Series FTIR spectrometers use a helium-neon laser operating in the visible region at 632.8 nanometers. The spectrometer is a Class 2 laser product, powerful enough to warrant caution in its use. Agilent Cary 600 Series FTIR spectrometers and microscopes comply with FDA and CE standards for light emitting products.

An attenuated portion of the laser beam passes into and through the spectrometer sample compartment. Although not powerful enough to harm your skin should your hand intercept it, the laser light could cause retinal (eye) damage during prolonged direct viewing. This is not possible given the normal optical layout of the spectrometer. However, if a highly reflective surface such as a mirror is allowed to intercept the beam, the beam could be redirected out of the sample compartment resulting in on-axis or direct viewing. Care must be taken to avoid this.

The laser in the spectrometer is operating when the green power indicator of the spectrometer is active. The Agilent Cary 600 Series FTIR spectrometers incorporate an interlock switch that automatically turns off power to the laser if the interferometer compartment cover is opened.

No maintenance of the spectrometer or microscope by users is required to maintain specifications, proper operation, and compliance with FDA and CE standards for light-emitting products.

Labeling

Laser safety labels are located on the rear of the Agilent Cary 600 Series FTIR spectrometer and underneath the top cover.



Figure 1. Laser safety label locations on the rear of the Agilent Cary 600 Series FTIR spectrometer

The top left safety warning says ‘Class 2 laser radiation when open. Do not stare into beam’.

The IEC Warning and explanatory label is located on the center top of the rear of the 600 Series FTIR spectrometer. The laser safety warning states ‘Laser Light. Do not stare into beam. Class 2 laser product 600 mW, 632.8 nm CW EN60825-1:2007’.

Rear of instrument




Figure 2. Laser safety labels on the rear of the Agilent Cary 600 Series FTIR spectrometer

Under top cover

The following image shows the laser safety warning located on the top of, and underneath, the interferometer compartment cover.




Figure 3. Laser safety warning located on top of internal compartment cover

Near apertures

The following image shows the laser safety warning located near all apertures on both the Agilent Cary 600 FTIR Series instruments and microscopes.



Figure 4. Laser safety warning located near apertures

Other precautions

Infrared sources operate at high temperatures, which may burn you. Before replacing a source element that has been lit, switch off the instrument and wait for the source to cool.

Do not block the ventilation grills on the instrument and accessories. Consult the manual(s) supplied with your computer, monitor and printer for their specific ventilation requirements.

Use of the FTIR spectrometer and the Oil Analysis accessory may involve materials, solvents and solutions which are flammable, corrosive, toxic or otherwise hazardous.

Careless, improper, or unskilled use of such materials, solvents and solutions can create explosion hazards, fire hazards, toxicity and other hazards which can result in death, serious personal injury, and damage to equipment and property.

Always ensure that laboratory safety practices governing the use, handling and disposal of such materials are strictly observed. These safety practices should include the wearing of appropriate safety clothing and safety glasses.

Use only spare parts from Agilent with your instrument.

Warning and caution messages

Carefully read all warnings and cautions and observe them at all times.

A Warning message is used in the text when failure to observe instructions or precautions could result in death or injury. Warnings have the following format:

WARNING



Hazard Type

Nature of the hazard, information on how to avoid the hazard, and possible consequences if you don't.

The triangular symbols that appear in conjunction with warnings are outlined in the next section.

A Caution message is used when failure to observe instructions could result in damage to equipment (Agilent-supplied and/or other associated equipment). Cautions have the following format:

CAUTION

Caution information appears here.

Information symbols

The following triangular symbols appear in conjunction with warnings on the spectrometer and associated documentation. The hazard they depict is shown below each symbol:



Corrosive liquid



Electrical shock



Eye hazard



*Heavy weight
(danger to feet)*



*Heavy weight
(danger to hands)*



Hot surface



Laser hazard



Moving parts










Noxious gas



The following symbol may be used on warning labels attached to the instrument. When you see this symbol, refer to the relevant operation or service manual for the correct procedure referred to by that warning label.

The following symbols also appear on the instrument or in the documentation:

	Mains power on.
	Mains power off.
	Fuse.
	Single phase alternating current.
	Direct current.
	When attached to the rear of the instrument, indicates that the product complies with the requirements of one or more EU directives.
	When attached to the rear of the product, indicates that the product has been certified (evaluated) to CSA 61010.1 and UL 61010-1.

Color coding

The various indicator lights appearing on Agilent instruments and associated accessories are color-coded to represent the status of the instrument or accessory.

- A green light indicates the instrument is in normal or standby mode.
- A yellow light indicates that the instrument needs attention.
- A blue light indicates that operator intervention is required.

CE compliance

Agilent Cary FTIR spectrometers and the Agilent Oil Analysis accessory have been designed to comply with the requirements of the Electromagnetic Compatibility (EMC) Directive and the Low Voltage (electrical safety) Directive (commonly referred to as the LVD) of the European Union. Agilent has confirmed that each product complies with the relevant directives by testing a prototype against the prescribed EN (European Norm) standards.

Proof that a product complies with the directives is indicated by:

- The CE marking appearing on the rear of the product.
- The documentation package that accompanies the product, containing a copy of the Declaration of Conformity. This declaration is the legal declaration by Agilent that the product complies with the directives and also shows the EN standards to which the product was tested to demonstrate compliance.

Electromagnetic Compatibility

EN55011/CISPR11

Group 1 ISM equipment: group 1 contains all ISM equipment in which there is intentionally generated and/or used conductively coupled radio-frequency energy which is necessary for the internal functioning of the equipment itself.

Class A equipment is equipment suitable for use in all establishments other than domestic and those directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.

This device complies with the requirements of CISPR11, Group 1, Class A as radiation professional equipment. Therefore, there may be potential difficulties in ensuring electromagnetic compatibility in other environments, due to conducted as well as radiated disturbances.

Operation is subject to the following two conditions:

- 1** This device may not cause harmful interference.
- 2** This device must accept any interference received, including interference that may cause undesired operation.

If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try one or more of the following measures:


- 1** Relocate the radio or antenna.
- 2** Move the device away from the radio or television.
- 3** Plug the device into a different electrical outlet, so that the device and the radio or television are on separate electrical circuits.
- 4** Make sure that all peripheral devices are also certified.
- 5** Make sure that appropriate cables are used to connect the device to peripheral equipment.
- 6** Consult your equipment dealer, Agilent Technologies, or an experienced technician for assistance.
- 7** Changes or modifications not expressly approved by Agilent Technologies could void the user's authority to operate the equipment.

ICES/NMB-001

This ISM device complies with Canadian ICES- 001.

Cet appareil ISM est conforme à la norme NMB-001 du Canada.

This page is intentionally left blank.



2. Introduction

Specifications	21
Installation requirements	24
Training	26
Documentation	27

This manual contains the information required to successfully get started using the Agilent Cary FTIR Oil Analyzer system, or the Oil Analysis accessory with your FTIR spectrometer.

This manual is included with the:

- Agilent Cary 600 Series FTIR Oil Analyzer
- Agilent Oil Analysis accessory

Unless mentioned, the specifications of the Agilent Cary FTIR Oil Analyzer system and the Oil Analysis accessory installed on an existing Agilent FTIR spectrometer are the same.

Additional information on the optional Advanced Autosampler is included in Chapter 5.

An FTIR Oil Analyzer system, whether delivered as a bundled system or where an Oil Analysis accessory is to be installed on an existing FTIR spectrometer, comprises:

- An Agilent Cary FTIR spectrometer (bundled system includes an Agilent Cary 640 FTIR spectrometer)
- The Oil Analysis accessory (includes Oil Analyzer software)
- Advanced Autosampler (optional for a fully-automated system)
- A computer and its peripherals (see Page 26 for requirements)

The FTIR Oil Analyzer is a system optimized for sampling and analyzing synthetic and petroleum-based lubricant and hydraulic fluids. The system is designed to meet the needs of the U.S. Department of Defense Joint Oil Analysis Program (JOAP) as well as commercial applications.

The Agilent Cary FTIR Oil Analyzer is controlled by a computer running the Oil Analyzer software, which is a Resolutions Pro interface dedicated to controlling an FTIR spectrometer fitted with an Oil Analysis accessory and the optional Advanced Autosampler.

Sample fluids are pumped through a filter and tubing outside the optics and electronics area of the spectrometer and then through a zinc selenide (ZnSe) flow-through cell. The Oil Analysis accessory mounts the cell within the sample compartment of the spectrometer, where the fluid is analyzed.

The following sequence describes how samples are measured and analyzed using an FTIR Oil Analyzer system:

- 1** A background spectrum of an empty, clean sample cell is collected.
- 2** The sample probe is inserted into a bottle containing the oil to be analyzed. This is achieved manually for manual and semi-automated analyses but the optional Advanced Autosampler facilitates automated analysis of multiple samples.
- 3** A peristaltic pump pulls the sample through a 38- μm (400-mesh) filter element and then into the 100- μm pathlength, flow-through cell.
- 4** The infrared absorbance spectrum of the sample is collected and interpreted by the computer.
- 5** A report comparing the sample spectrum with JOAP standards is generated.
- 6** The peristaltic pump is used to remove the sample from flow-through cell and then solvent is drawn through the sample flow path to clean the system.

If you have purchased an Agilent Cary 600 Series FTIR Oil Analyzer, refer to the User's Guide that shipped with your Agilent Cary 600 Series FTIR spectrometer for additional information.

If you have purchased the Agilent Oil Analysis accessory only, skip all instructions related to the initial setting up of your Agilent spectrometer. Instead, refer to the manual(s) that shipped with your spectrometer for additional information.

Specifications

The Agilent Cary FTIR Oil Analyzer system is suitable for indoor use **only** and is classified Pollution degree 2 and Installation Category II (EN 61010-1).

Environmental

The conditions required for proper operation of your Agilent Cary FTIR Oil Analyzer are:

- Temperature: 20 °C to 26 °C (68 °F to 80 °F)
- Temperature gradient: 1.0 °C/h (1.8 °F/h) maximum
- Relative humidity: 20% to 50% (non-condensing)
- Altitude: 3000 m (10000 ft) maximum
- Free from corrosive and flammable fumes
- Free from strong electromagnetic fields
- Free from vibrations

Performance degradation may result from the exposure of the spectrometer to strong radio frequency energy. If degradation is experienced, you should reorient or relocate the spectrometer or the radio frequency source.

WARNING**Heavy Weight Hazard**

Danger to hands and feet. Incorrect handling of the spectrometer can lead to injury and/or damage to the instrument. The Agilent Cary 600 Series FTIR spectrometers weigh approximately 80 kg (176 lb). Always use a suitable lifting device and proper lifting techniques when moving the instrument. Do not attempt to lift the spectrometer unassisted.

For more information or assistance, contact your Agilent representative.

Power

The power specifications for your Agilent Cary FTIR Oil Analyzer are:

- Spectrometer voltage requirement: 100–240 VAC
 $\pm 1\%$; 50-60 Hz ± 1 Hz
- Spectrometer power consumption: 200 VA (286 W)
- Spectrometer heat dissipation: 500 BTU/h (147 W)
- Peristaltic pumps: 3.15 A at 115 VAC; 1.6 A at 230 VAC
- Advanced Autosampler (optional): 4 A maximum at
100-240 VAC; 47-63 Hz
- Dedicated, clean 10-A power circuit with 3-pin earth-grounded
mains socket-outlets.
- No other instruments, machinery or appliances should be
powered from the same circuit.

- Each of the system components – FTIR spectrometer, peristaltic pump(s), Advanced Autosampler (optional), computer, monitor, and printer – requires an earth-grounded mains socket-outlet. It is useful to have at least two additional sockets available for additional peripheral devices or spectrometer accessories. Depending on the computer options, monitor, peripherals, and spectrometer accessories chosen, it is possible that a second dedicated clean power circuit be required.
- See the vendor's documentation that came with the computer, monitor, and printer for the power requirements of each device.
- Fuse information on the rear of the instrument is the most up-to-date. See the Agilent Cary 600 Series Spectrometer User's Guide for fuse information.

Laser

Specifications for the laser light that is accessible during operation of your Agilent Cary FTIR spectrometer are:

- Maximum accessible power: < 600 μ W
- Pulse duration: Continuous
- Wavelength: 632.8 nm
- Maximum beam divergence: Cone of angle 15 degrees or less coming to a focus in the center of the sample compartment and diverging thereafter.

Specifications for the embedded laser in the spectrometers are:

- Maximum power output: < 5 mW
- Pulse duration: Continuous
- Wavelength: 632.8 nm
- Maximum beam divergence: 1.7 milliradians

Weights and dimensions

Table 1. Agilent Cary 600 Series FTIR Oil Analyzer weights and dimensions

Unit	Width (cm, in)	Depth (cm, in)	Height (cm, in)	Weight (kg, lb)
FTIR spectrometer	75, 31	70, 28	36.5, 15	75, 165
Peristaltic pump	18, 7.25	35, 14	18, 7.25	N/A
Advanced Autosampler	43, 17	58.5, 23	55, 21.5	N/A

WARNING



Heavy Weight Hazard

Danger to hands and feet. Incorrect handling of the spectrometer can lead to injury and/or damage to the instrument. The Agilent Cary 600 Series FTIR spectrometers weigh approximately 80 kg (176 lb). Always use a suitable lifting device and proper lifting techniques when moving the instrument. Do not attempt to lift the spectrometer unassisted.

CAUTION

You will need to allow for a further 10 cm (4 in) behind the rear panel of the spectrometer for cable connections and a further 70 cm (28 in) overhead for service access.

Installation requirements

Before receiving an Agilent Cary 600 Series FTIR Oil Analyzer, you will have received the Site Preparation Guide for the Agilent Cary 600 Series FTIR spectrometers, which describes the environmental and operating requirements of the system. You must prepare your laboratory according to these instructions before the FTIR Oil Analyzer system can be installed.

You should keep the Site Preparation Guide for future reference. If you have misplaced your copy, you can obtain a replacement from your local Agilent office.

Installation of an Agilent Oil Analysis accessory on an existing Agilent Cary FTIR spectrometer may require reorientation of your instrument by 90 degrees, such that the sample compartment faces to the side when you view the spectrometer from in front of the bench. This rotation ensures that the stand used to position the sample probe is located away from the edge of the worktable, as shown in Chapter 3, and that the (optional) Advanced Autosampler can be plumbed in correctly, see Chapter 5. Consult your Agilent representative for further information.

Laboratory

Before installation of your Agilent Cary FTIR Oil Analyzer system can proceed, verify that:

- Independent 10-amp power lines with grounded outlets for the FTIR spectrometer, peristaltic pump(s), Advanced Autosampler (optional), computer, monitor, and printer are available, and that the outlets are located within reach of the power leads of these devices.
- You have consulted your Agilent representative for power requirements for the system, and complied explicitly with the instruction provided.
- You have provided a level flat surface which is free of vibration and rigid enough to support the instrument and accessories without warping or sagging. Loaded flatness tolerance: 0.4 millimeters per 300 millimeters of length (1/64 inch per foot of length).
- There will be at least 10 centimeters (4 inches) of working space behind the rear panel of the FTIR spectrometer to afford access to the connectors.
- The computer will be no more than 2 meters (6 feet) from the instrument (the USB cable supplied by Agilent is 2 meters long).
- 30 centimeters (12 inches) will be available for the mouse (usually to the right of the computer).

Computer

Before installing your instrument, verify that your computer system meets the **minimum** requirements detailed in this section.

The recommended configuration should be followed when buying a new computer and is as follows:

- IBM compatible
- Intel® Pentium® 4 processor (2.6 GHz or greater)
- 512 MB of RAM
- 40-GB (or greater) hard drive
- USB 2.0 High Speed interface
- Video card supporting 1024 × 768 resolution, High Color (16-bit) mode
- Super VGA screen
- 24 x CD drive
- 16-bit sound card
- Windows® compatible keyboard
- Microsoft® or compatible mouse
- Microsoft Windows XP with Service Pack 2 using Administrator mode only

For the most recent list of minimum requirements, see the Agilent, Web site, www.agilent.com. If you are unsure of the requirements, consult with your Agilent representative to confirm that you have the correct specifications required for the Resolutions Pro software to correctly operate.

Training

The Resolutions Pro software, which is provided with the Oil Analysis accessory, is accompanied by a Help system that provides detailed information on the FTIR Oil Analyzer system.

If Agilent installs the instrument(s), the Agilent representative will demonstrate the basic operating procedures while conducting the installation performance tests during the installation procedure. The Agilent representative, however, is not necessarily experienced in complex analytical routines and is not authorized to conduct extensive training.

To ensure that your operators benefit the most from witnessing the installation performance tests, operator training should be completed before your equipment is installed. It is strongly recommended that you take advantage of the special training courses that are conducted at various locations by the Agilent customer support and sales organization. In some areas, it may be possible to arrange for operator training to be carried out after the installation, using your own instrument. To investigate this possibility, contact your local Agilent sales and service office.

Documentation

This User's Guide provides basic instructions for installing, operating and maintaining your Agilent FTIR spectrometer with an Oil Analysis accessory.

NOTE

Where procedures differ for the Agilent Cary FTIR Oil Analyzer system and the Oil Analysis accessory, they will be clearly identified in this manual.

Chapter 5 of this manual contains information on how to set up and use the optional Agilent Advanced Autosampler.

If you have purchased an FTIR Oil Analyzer system, refer to the User's Guide that accompanied your Agilent 600 Series FTIR spectrometer.

Further information and instruction for the Resolutions Pro software, analysis methodology, operating procedures, and details and instruction for various accessories is provided in the Help and electronic manuals loaded onto your computer hard drive during software installation.

Conventions

The following conventions have been used in procedures throughout the documentation:

- Bolded items indicate an action. For example, 'click **OK**' and 'From the **Edit** menu, choose **Copy**'.
- ALL CAPITALS indicate keyboard commands. For example, 'press ENTER' and 'press SHIFT+F8'.

Throughout this document:

The Cary 600 Series FTIR refers to any of these Agilent FTIR spectrometers:

- Agilent Cary 640 FTIR
- Agilent Cary 660 FTIR
- Agilent Cary 670 FTIR
- Agilent Cary 680 FTIR

The Scimitar Series refers to any of these Agilent FTIR spectrometers:

- Agilent 800 FTIR
- Agilent 1000 FTIR
- Agilent 2000 FTIR

The Excalibur Series refers to any of these Agilent FTIR spectrometers:

- Agilent 3100 FTIR
- Agilent 3600 FTIR
- Agilent 4100 FTIR

An Agilent Cary FTIR Oil Analyzer refers to a system comprising of:

- An Agilent Cary FTIR spectrometer
- Agilent Oil Analysis accessory (includes Oil Analyzer software)
- Agilent Advanced Autosampler (optional)
- A computer and its peripherals



3. Unpacking and Installation

Unpacking	29
Installation space requirement	31
Setting up	31
Installing Resolutions Pro	40

This chapter includes information on the hardware and installation of the Agilent Cary FTIR Oil Analyzer system. It is applicable equally to the system purchased as a bundle and the Oil Analysis accessory, which allows you to reconfigure your existing FTIR spectrometer.

NOTE

Chapter 5 of this manual contains information on how to set up and use the optional Agilent Advanced Autosampler.

Unpacking

As soon as the shipment arrives:

- 1 Locate the shipping list documentation and check that you have received all of the items listed.

NOTE

If any of the listed items are missing, stop immediately and call Agilent customer service.

- 2 Carefully check the exterior of the shipping containers for any external damage. (Stains on the containers may indicate exposure to water.)

CAUTION

Do not open the shipping containers at this time.

- 3 If the shipping container(s) appear to be damaged (water damage, crushed package, and so on), contact Agilent and the carrier within five days. Do not accept packages with obvious puncture damage. All other damage should be noted on the shipping document and signed by the delivery agent.
 - 4 If you are sure that all boxes have been delivered and appear to be undamaged, place them where their contents can come to room temperature.
-

WARNING



Heavy Weight Hazard

Danger to hands and feet. Incorrect handling of the spectrometer can lead to injury and/or damage to the instrument. The Agilent 600 Series FTIR spectrometers weigh approximately 80 kg (176 lb). Always use a suitable lifting device and proper lifting techniques when moving the instrument. Do not attempt to lift the spectrometer unassisted.

CAUTION

Do not unpack any additional boxes or packages or the spectrometer or microscope. Your Agilent representative will complete the unpacking and installation of your Agilent Cary FTIR spectrometer. If unpacked incorrectly, equipment may become damaged.

Installation space requirement

You must provide bench space of at least 150 centimeters across by 90 centimeters deep (5 feet by 3 feet) to hold the FTIR Oil Analyzer system including the computer, with enough room to use the keyboard and mouse.




Figure 5. Space required for an Agilent Cary FTIR Oil Analyzer system

Setting up

This section describes how to physically set up an Agilent Oil Analysis accessory on your FTIR spectrometer.

Spectrometer

This section describes the procedure for setting up your FTIR spectrometer for use as an FTIR Oil Analyzer. Without the optional Advanced Autosampler, the spectrometer and accessory are arranged as illustrated in Figure 6.




Figure 6. External view of an Agilent Cary FTIR Oil Analyzer system.

A: Oil Analyzer B: Sample cell C: Sample pump
D: Sample probe E: Sample F: Cell drain line

To set up the spectrometer for use as an FTIR Oil Analyzer:

- 1 Put the FTIR spectrometer to the left side of the assigned space, with its sample compartment oriented to the right as you are facing it from the bench.

WARNING



Heavy Weight Hazard

Danger to hands and feet. Incorrect handling of the spectrometer can lead to injury and/or damage to the instrument. The Agilent 600 Series FTIR spectrometers weigh approximately 80 kg (176 lb). Always use a suitable lifting device and proper lifting techniques when moving the instrument. Do not attempt to lift the spectrometer unassisted.

- 2 Remove the cover for the sample compartment and install the sample cell holder from the Oil Analysis accessory.

- 3 Slide the base of the C-shaped pump stand under the spectrometer. The stand should be positioned such that the pump head protrudes over the sample compartment of the spectrometer, as shown in Figure 7.
- 4 Position the stand for the sample probe on the bench, in front of the sample compartment.



Figure 7. Sample cell and pumping components of the FTIR Oil Analyzer
A: Sample pump B: Filter C: Sample cell D: Sample probe

Pump

The main plumbing components of the FTIR Oil Analyzer important during routine use are shown in Figure 7, which depicts a front view of the system without an Advanced Autosampler.

Unpacking and Installation

The sample pump (A) is placed on the upper level of a C-shaped pump stand, which is designed so the lower level slides underneath the base of the spectrometer. The pump head is positioned above the sample compartment of the spectrometer.

Figure 8 depicts a front view of the peristaltic pump. The speed control is (A) located on the upper right front corner of the pump.

NOTE

For manual operation, adjust the vernier scale of the speed control to '8.00' and then lock it in place.




Figure 8. Peristaltic pump – front view

A: Speed control B: Pump tubing release lever C: Power and direction switch

Located below the speed control, the three-position switch (C) determines whether the pump is active (or not) and the direction in which the peristaltic flow is induced.

Table 2. Function of each position of the switch on the sample pump

Position	Function
Left (counterclockwise)	Pump pumps sample or solvent through the filter and into the flow-through cell.
Center	Pump is off.
Right (clockwise)	Pump draws fluid from the sample cell back through the filter and to the probe.

NOTE

Orientations referred to in Table 2 assume you are facing the front of the pump.

Using the release lever (B), the upper part of the pump head can be raised to remove tubing. The pump head is raised typically if:

- The FTIR Oil Analyzer system will not be used for a while
- Tubing requires replacement

WARNING**Moving Parts Hazard**

Danger to hands. To avoid exposing moving parts, always operate your peristaltic pump with the tubing installed and the pump head lowered. While the peristaltic pump is running, do not put your hands on, or near, the pump head.

NOTE

During routine operation of your system, the three-position switch is the only component of the pump that you need to manipulate. All the measuring, monitoring, system checks, and report generation functions of the FTIR Oil Analyzer are automatically controlled by the Oil Analyzer software.

Sample probe

Liquid (either a sample or wash solvent) is taken up by the sample probe and it then passes through the filter assembly into the flow-through cell. When performing oil analysis manually or in semi-automated mode the sample probe is held in position by a stand that can accommodate differently sized sample containers.

WARNING

Eye Hazard

Samples and solvents used when performing oil analysis are hazardous and pose a danger to eyes. Always ensure you wear appropriate personal protective equipment.

Zinc selenide sample cell

The zinc selenide (ZnSe) sample cell has a flow-through design and a transmission pathlength of 100 micrometers. The stainless-steel-protected cell is placed into the sample-cell holder, which mounts within the sample compartment as illustrated in Figure 10. The input and output connections are made to the flow-through cell using plastic tubing.

WARNING

Noxious Gas Hazard

Zinc selenide reacts with acids to form toxic hydrogen selenide gas (H_2Se). Do not use acid during sample preparation of samples that are to be introduced to ZnSe cells.



Figure 9. Zinc selenide flow-through cell



Figure 10. Flow-through cell located in cell holder

The peristaltic action, which is induced by the sample pump, draws fluid to be analyzed through a filter assembly and then into the flow-through cell. From the output of the sample cell, any waste fluid is directed to a bottle. Alternatively, by reversing the pump direction, fluid can be pumped back through the filter assembly to a waste bottle at the sample probe location.

Sampling plumbing

This section describes how to set up and check the sampling plumbing for your system. When operating the Oil Analysis accessory pulls samples through a metal mesh filter and then the ZnSe sample cell, finally to a waste container.

To plumb the Oil Analysis accessory:

- 1 Locate the ZnSe sample cell. The interior of this flow-through cell should appear clear, with a yellow-gold color. Check the cell for any cracks in the windows.
- 2 Slide the sample cell onto the guides located in the sampling slot on the sample-cell holder. The inlet and outlet tubes should be in front of the cell as it is placed onto the guides. Gently move the sample cell up and down whilst sliding it into the slot, until the ball lock engages. The cell should be completely in the sample-cell holder, with only the black metal tab and the inlet and outlet tubes outside.
- 3 Ensure that the peristaltic pump is switched off, by making sure its switch is in the center position.
- 4 Rotate the pump tubing release lever fully counterclockwise to raise the pump head.
- 5 Route the sample tubing, which connects the inlet of the flow-through cell to sample probe, through the pump head. Loop the yellow, Tygon tubing up and over the rollers, so that the probe end is to the right of the pump head.

NOTE

You may need to lift the black, spring-loaded retainers while routing the Tygon tubing through the pump head.



Figure 11. Sample pump with pump head raised and tubing in place

- 6** Holding tension on the two sides of the tubing in the pump head, rotate the pump tubing release lever fully clockwise to close the pump head, until it latches.

NOTE

Before lowering the pump head, ensure that the Tygon tubing is centered correctly between the retainers.

- 7** Connected to the outlet of the sample cell, the ends of both lengths of Tygon tubing should be placed into the opening of an approved waste container.
- 8** Position the ends of the tubings above the level of the used oils and solvents that will be collected in the waste container, as lower would allow effluent to be pulled back into the flow-through cell when reversing the pump.

CAUTION

To avoid system contamination, the waste container should be located > 30 cm (12 in) below the level of the bench on which the FTIR Oil Analyzer is supported.

You are now ready to check the sampling system plumbing.

To check the plumbing of the Oil Analysis accessory:

- 1** Ensure that the peristaltic pump is connected to the power supply.
- 2** Move the switch on the pump housing to the left position. The pump drive head should start rotating counterclockwise.
- 3** Immerse the end of the sample probe in a small amount of wash solvent, either heptane or hexane, in a beaker or similar container. Verify that the liquid is pulled up the sample tubing, through the filter, and into the sample cell.
- 4** Keep the beaker/container under the sampling probe and move the three-position switch to the far right position. Confirm that the wash solvent is pumped from the sample cell back into the beaker.
- 5** Remove the end of the sample probe from the liquid.
- 6** Move the switch back to the left position and allow air to be pulled through the system for about 15 seconds.
- 7** Move the switch to its center position to stop the sample pump.

The sample plumbing is now ready for operation.

Installing Resolutions Pro

Agilent Oil Analyzer software is the interface that you need to use to control an FTIR spectrometer fitted with the Oil Analysis accessory. This software will be installed by your Agilent representative at the same time as the Resolutions Pro software that is required to control your FTIR spectrometer.

Software

Your Agilent representative will install the Resolutions Pro software during the process of installing the FTIR Oil Analyzer. However, should you need to re-install the software (for example, if the computer has been replaced), instructions are provided on how you install Agilent Resolutions Pro software.

NOTE

The computer must meet the minimum requirements stated previously on Page 26. For the most recent list of minimum requirements see the Agilent website, www.agilent.com. If you are unsure of the requirements, consult your Agilent representative to confirm that you have the correct specifications required for the Resolutions Pro software to correctly operate.

If you purchased the FTIR Oil Analyzer as a system, refer to the Cary 600 Series FTIR User's Guide for the software installation procedure.

CAUTION

If you purchased the Oil Analysis accessory and you need to upgrade from Resolutions Pro version 4.1 to version 5.0, follow the instructions provided with the upgrade disk.

CAUTION

Before installing or upgrading Resolutions Pro software, ensure your spectrometer is not connected to the computer.

Upon completing the installation, the USB cable to your spectrometer will be connected to the computer and your FTIR Oil Analyzer is ready to be used. With the software installed you need to install the Resolutions Pro Help and the sample data, both of which are provided on a separate disk.

NOTE

There is a 60-day trial period to use the Resolutions Pro software. You must register your software with Agilent before the trial period has ended. Details on how to do this are found in the installation instructions previously mentioned.

Help

The Help for Resolutions Pro software is provided on a separate disk. Resolutions Pro Help is installed on your computer after the Resolutions Pro software installation.

In addition to content on the Agilent Cary FTIR Oil Analyzer system, the Resolutions Pro Help includes information and instructions on:

- Safety
- Hardware
- Software menus and document types
- How to perform a collect, process and manipulate the data
- Create scripts to automate data collection, processing and manipulation
- How to install hardware accessories and components
- Maintenance and troubleshooting

For further information on the installation procedure, refer to the Resolutions Pro Help section of your Cary 600 Series FTIR User's Guide or the instructions provided with the upgrade disk.

Click the 'Help' button in Oil Analyzer software to display information that relates to the Oil Analyzer software, which contains links to additional content on the FTIR Oil Analyzer system and the Oil Analysis accessory. Alternatively, and as detailed within the next section, you can access Resolutions Pro Help through the Resolutions Pro software. Browse for information of interest to you using the index, locate specific pages titles via the contents list, or perform keyword searches.



4. Operation

Power connections	44
Spectrometer warm-up	46
Loading samples	47
Cleaning the cell	48
Starting the software	49
Turning off the system	50

The following procedures for the Agilent Oil Analyzer system assume that the Agilent Oil Analysis accessory and accompanying spectrometer have already been set up in your location with the Oil Analyzer software installed.

For additional hardware information, see the Resolutions Pro Help.

To access the FTIR Oil Analyzer section of the Help:

- 1 Start the Oil Analyzer software.
- 2 Click **Help** on the toolbar to display the Help.
- 3 Click the **Contents** tab if necessary.
- 4 Click the plus (+) icon next to 'About Your FTIR Systems' to expand it.
- 5 Click the plus (+) icon next to 'Agilent 600 FTIR Series' to expand it for hardware information on your spectrometer, or click the plus icon next to 'FTIR Oil Analyzer' for hardware information on your system.

Alternatively:

- 1 Start the Resolutions Pro software.
- 2 Press F1 on the keyboard to display the Help.
- 3 Follow Steps 3 to 5 above.

NOTE

If you have purchased an Agilent Cary FTIR Oil Analyzer, see the User's Guide supplied with your 600 Series FTIR spectrometer. If you have purchased only the Agilent Oil Analysis accessory, refer to the manual(s) supplied with your existing FTIR spectrometer.

Power connections

This section details the correct procedures to connect power cords to the components of your FTIR Oil Analyzer system.

Spectrometer

Power to your Agilent FTIR spectrometer is connected on the rear panel of the instrument.

To connect power to your spectrometer:

- 1 Make sure the power switch on the rear panel of the spectrometer is in the off (0) position.




Figure 12. Power switch and socket on the rear panel of the spectrometer

- 2 Connect the power cord to the power socket located underneath the power switch on the rear panel of the spectrometer.
- 3 Connect the other plug of the power cable to the electrical power outlet.

Lights

There are two LEDs on the front of a Cary 600 Series FTIR spectrometer in the lower left corner. Table 3 summarizes the LED behavior.

Table 3. LED descriptions

Position	Condition	Comment
Top — Power/Scan	Flashing yellow	The spectrometer is initializing.
	Solid yellow	The spectrometer has not correctly initialized and needs attention.
	Solid green	The spectrometer has initialized and is ready.
	Flashing green	The spectrometer is scanning.
Bottom — A/D over-range	Off	Normal operation.
	Flashing blue	The A/D exceeds its maximum value. Reduce the sensitivity setting or increase the beam attenuation.

Peristaltic pump(s)

Power is connected to the peristaltic pump(s) on the rear panel of the pump.

To connect power to a peristaltic pump:

- 1 Make sure the power switch on front of the pump is in the off (0) position.
- 2 Connect the power cord to the socket at the rear of the pump.
- 3 Connect the other end of the power cable to the electrical power outlet.

Spectrometer warm-up

If the FTIR Oil Analyzer is set up but the system has been turned off, you will need to turn on the spectrometer and allow it sufficient time to warm up.

The lifetime of the electrical and optical components can be extended by keeping the spectrometer on and warm. It is unnecessary to turn the spectrometer off overnight or when it is not being used, unless this will be for an extended period.

To reinitialize an FTIR spectrometer that has been turned off:

- 1 Turn on the power switch located on the back of the spectrometer.
- 2 Wait at least 30 minutes for the infrared source in the spectrometer to reach its operating temperature.

NOTE

If the spectrometer has been turned off for an extended period, any data you collect before the spectrometer has reached its operating temperature may not be accurate.

-
- 3 Turn on the computer and monitor, wait for the operating system to load.
 - 4 Connect the USB cable from your spectrometer to the computer. The new hardware should be recognized by the drivers installed during the installation of Resolutions Pro software.
 - 5 If you are setting up your FTIR Analyzer accessory or if you change the spectrometer detector, beamsplitter, or source, you must configure the hardware before you can collect any data using your system. If your spectrometer is already configured as an FTIR Oil Analyzer, you are ready to start the Oil Analyzer software.

Loading samples

A peristaltic pump is used to draw either fluids (oil or wash solvent) or air through the flow-through cell of the Oil Analysis accessory, which is mounted in a cell holder within the sample compartment of the spectrometer. There are two ways that the sample pump of the FTIR Analyzer can be operated:

- Continuous pumping in manual or semi-automated mode
- Pulse pumping (automated sampling using the Advanced Autosampler only)

In continuous pumping, an oil sample is drawn through the sample probe and into the flow-through cell using the peristaltic pump. The Oil Analyzer software monitors the absorption of light by the contents of the sample cell and uses that measurement to determine if the cell is completely full. Sample data is not acquired unless the application is satisfied that the flow-through cell is full, which is defined within the administrator settings.

Pulse pumping is particularly useful for automated applications that analyze low-viscosity oils or when there is only a small amount of sample available for analysis, which when using continuous pumping could result in the complete sample being drawn through the sample cell before the analyzer is ready to collect data. Using pulse pumping, an automated method transports an aliquot of oil with air gaps either side. The size of the aliquot is the result of pre-determined timings in the lowering and raising of the sample probe by the Advanced Autosampler.

NOTE

When using the Advanced Autosampler (optional) there are two pumps in operation: the sample pump, which delivers oil samples to the flow-through cell; and a wash pump, which rinses the sample probe and automates cell cleaning.

Cleaning the cell

Depending on your hardware and the acquisition mode that you are operating the system, cleaning the sample cell involves either the manual or automatic pumping of wash solvent from a reservoir into and through the sample cell. Air is then drawn through the cell to remove residual solvent.

NOTE

The Oil Analyzer software monitors the cleanliness of the sample cell and will continue pumping wash solvent into the cell until it is sufficiently clean. The cleanliness standard for the sample cell is defined within the administrator settings.

For additional information on cleaning your ZnSe sample cell and either of the pumping modes used for sampling, see the Resolutions Pro Help.

To access the Troubleshooting section of the Help:

- 1 Start the Oil Analyzer software.
- 2 Click **Help** on the toolbar to display the Help.
- 3 Click the **Contents** tab if necessary.
- 4 Click the 'plus' (+) icon next to 'About Your FTIR Systems' to expand it.
- 5 Click the 'plus' (+) icon next to 'FTIR Oil Analyzer' for hardware information on your system.

Alternatively:

- 1 Start the Resolutions Pro software.
- 2 Press F1 on the keyboard to display the Help.
- 3 Follow Steps 3 to 5 above.

Starting the software

To start the Oil Analyzer software click 'Start' > 'Programs' > 'Agilent Resolutions' > 'Oil Analyzer'.

If you are doing the same type of analysis as the most recent run and you do not need to modify any of acquisition or processing of data, you may proceed to analyze your oil samples. Otherwise, if this is the first analysis since your FTIR Oil Analyzer was installed or you need to modify the acquisition or data processing parameters, you need to login to the Oil Analyzer software as an administrator to configure your system.

To log on as an administrator:

- 1 From the Oil Analyzer software main window, click **Administration**. The password dialog window will appear.
- 2 Enter your password. Asterisks will appear as you type. Click **OK**.

NOTE

The default password set up during installation is 'Agilent', typed without the single quotation marks.

- 3 You are now logged in as an administrator. The 'General Options' dialog box will appear.

See the Help for detailed information on the Oil Analyzer software.

To access the Oil Analyzer Software section of the Help:

- 1 Start the Oil Analyzer software.
- 2 Click **Help** on the toolbar to display the Help.

Alternatively:

- 1 Start the Resolutions Pro software.
- 2 Press F1 on the keyboard to display the Help
- 3 Click the **Contents** tab if necessary.
- 4 Click the 'plus' (+) icon next to 'Software' to expand it.
- 5 Click **Oil Analyzer Software**.

Turning off the system

Unless the system will be idle for some time or will be moved, Agilent recommends that the spectrometer be left turned on. The peristaltic pump(s), autosampler (if you have one) and the computer (along with its peripherals) can be turned off.

NOTE

Agilent recommends that you disable the screen saver because it may interfere with long data collections.

To shut down the computer:

- 1 Click **Exit** to close Oil Analyzer software.
- 2 Close all other applications.
- 3 Select **Shut down** from the **Start** menu and remember to switch off your monitor.


CAUTION

Using the computer's power switch to quit Windows may cause data to be lost. Always use the procedure advised by Microsoft.

If necessary, turn off the spectrometer.

NOTE

The lifetime of the tubing is improved by releasing the pump heads, which is achieved by rotating the release lever counterclockwise.



5. Advanced Autosampler

Installation space requirement	52
Setting up	53
Checking	56

This chapter describes how to:

- Connect the Agilent Advanced Autosampler to your Oil Analysis accessory.
- Confirm operation of your FTIR Oil Analyzer as a fully-Automated system.



Figure 13. FTIR Oil Analyzer with an Advanced Autosampler

The Advanced Autosampler consists of an X-Y-Z robotic arm that moves a sampling probe capable of accessing bottles held within a sampling space of 470 millimeters (X) by 200 millimeters (Y) by 115 millimeters (Z).

Sample bottles are held in individual, interchangeable racks and are pinned so that the rack can only be correctly positioned on the spill tray in one orientation. The spill tray is pinned so that it can only be orientated on the Advanced Autosampler in one position, which ensures reproducible bottle placement.

The Agilent Advanced Autosampler includes an integral, high efficiency wash station with a software-enabled option of returning excess wash solvent to a wash reservoir or to waste. Two computer-controlled, variable-speed (6 to 600 revolutions per minute) peristaltic pumps supply the samples and wash solvent to the FTIR Oil Analyzer system. Pump speeds can be adjusted to supply samples at maximum speed or at slower speeds to minimize sample-volume requirements and reduce solvent usage.

Installation space requirement

You must provide bench space of at least 220 centimeters across by 120 centimeters deep (7 feet by 4 feet) to hold the FTIR Oil Analyzer system including the computer, with enough room to use the keyboard and mouse.




Figure 14. Schematic view of FTIR Oil Analyzer with Autosampler

Setting up

This section describes how to set up the Advanced Autosampler to work with an FTIR Oil Analyzer system.

To connect your Advanced Autosampler to your FTIR Oil Analyzer system:

- 1 Ensure that your sample pump is turned off and that your computer is not running the Oil Analyzer software.
- 2 Place the additional peristaltic pump and the Advanced Autosampler to the right side of the spectrometer, as shown in Figure 14. The tray of the autosampler faces towards the sample compartment of the spectrometer.

NOTE

The 'wash pump' supplies wash solvent to the wash station located at the sample probe's 'Home' position. The wash pump is identical to the sample pump already on the system but is controlled separately by the Oil Analyzer software.

-
- 3 Remove the two screws holding the top cover on the back of the autosampler. Lift and remove the cover and remove any shipping restraint from the sample probe carriage.
 - 4 Replace the top cover. Ensure the wash pump and autosampler are switched off, and then plug in the power cables.

CAUTION

Do not turn on the power.

-
- 5 Using the supplied RS-232 cable, connect the COM 1 port on your computer to the autosampler port labeled 'Computer' (or 'AA Port').
 - 6 Plug the bifurcated, pump-control cable into the autosampler port labeled 'Auxiliary' (or 'Prep Station').
 - 7 Connect the DB-9 connectors labeled 'Sample' and 'Wash' to the sample and wash pumps respectively. Switch both of the pumps to the 'Remote' setting.

- 8 Place the spill tray on the autosampler, taking care that the tray fits on the locating pins.
- 9 Slide the metal sample probe through the head of the probe holder and locate it into the hole on the bottom guide.




Figure 15. Sample probe positioned too high.

NOTE

If the head of the sample probe is positioned too high for the probe to go through the bottom guide plate, the head should be carefully lowered in to position by hand.

- 10 When the sample probe is positioned correctly, that is, such that the end of the sample probe extends 3 to 4 millimeter (0.5 inch) below the bottom guide plate, tighten the knurled knobs on the holder to secure the probe in place.

CAUTION

The autosampler will jam if you do not have the probe located through the bottom guide plate.

- 11 Take the length of Tygon waste tubing and attach it to the lower barbed fitting on the wash station of the autosampler.

NOTE

This is the waste line for excess wash solvent. The yellow tubing can be routed to a waste container.

CAUTION

This waste line, like those attached to the flow-through cell, should be located into the mouth of the waste container but above the level of the liquid. Doing this will prevent waste from being pulled back into wash station when reversing the direction of the wash pump.

- 12** Take the supplied wash tubing assembly, terminated at one end by a hollow weighted fitting, and attach the Viton tubing to the top barbed fitting of the wash station. A spare clamp is supplied to do this.
- 13** Rotate the pump tubing release lever fully counterclockwise to raise the pump head. Route the Viton tubing through the left side of the pump head, around the rotor, and out the right side.
- 14** Holding tension on the two sides of the tubing in the pump head, rotate the pump tubing release lever fully clockwise to close the pump head, until it latches.
- 15** The weighted end of this tubing assembly should be placed into the wash-solvent container, so that it is kept beneath the level of the liquid.

CAUTION

To prevent siphoning, the reservoir of wash solvent should be in a container that is located > 30 cm (12 in) below the level of the bench on which the FTIR Oil Analyzer is supported.

- 16** Adjust the vernier scale of the speed control on the sample pump to '7.00' and that on the wash pump to '2.00'. Lock the dials in place.

NOTE

Depending on the viscosity of your oil samples, you may need to adjust these settings for optimal performance of the system.

That completes the setup procedure for the Advanced Autosampler. You may now proceed to check that it is correctly functioning.

Checking

This section describes the procedures for checking that your Advanced Autosampler is correctly operating and ready to use. The 'Autosampler Tests' dialog box allows an administrator to test the individual functions of the Autosampler: to ensure that the pump controls are operational, to verify rack definitions, and to run the autosampler through some preliminary checks.

Before proceeding with an 'Autosampler Test' you must:

- 1 Place one of your sample racks on the spill tray of the autosampler, checking that the pins seat properly in the index holes of the rack
- 2 Turn on the Autosampler. The power button is located on the left side of the rear panel.
- 3 Ensure that the pump heads on the pumps are lowered and locked in place, and that the switch on the rear panel of each pump is in the 'Remote' position.

WARNING



Moving Parts Hazard

Danger to hands. To avoid exposing moving parts, always operate your peristaltic pump with the tubing installed and the pump head lowered. While the peristaltic pump is running, do not put your hands on, or near, the pump head.

- 4 Switch both the sample and wash pumps on to the far left position, that is, so that the pump heads will rotate counterclockwise.

NOTE

You may confirm this by temporarily switching the pump to Local.

- 5 If it is not already running, turn on the computer and start the Oil Analysis software.
- 6 Log on as an Administrator.

You are now ready to check your Advanced Autosampler is functioning correctly by performing a number of tests:

- Raise and lower the sample probe.
- Send the sample probe to its 'home' position.
- Send the sample probe to bottle #X.
- Start and stop the wash/sample pump.

CAUTION

Be careful when performing the autosampler tests. Avoid damage to your autosampler, by ensuring that the appropriate rack type is selected and that the racks are empty or sample appropriately sized bottles are in place.

To check the performance of your Autosampler:

- 1 Click the arrow to the right of the 'Administration' button, and click **Autosampler Test**.
- 2 From the Autosampler Tests dialog box, you can perform a number of tests to confirm that the Advanced Autosampler is correctly operating.

See the Help for further information of the individual tests.

Bottle numbering convention

Figure 16 shows how sample bottles are assigned numbers within a rack. The example shown is of a 4 x 10-rack arrangement but the numbering system is the same for all racks, that is, the bottle closest to the 'Home' position of the sample probe is assigned position #1.




Figure 16. Bottle numbering convention for autosampler rack



6. Maintenance

Routine	60
Cleaning	61
Spare parts	61

This chapter includes information on how to access the Agilent Cary FTIR Oil Analyzer maintenance procedures that may be carried out by an operator. Any maintenance procedures not specifically mentioned should be carried out only by Agilent-trained, Agilent-qualified or Agilent-authorized representatives.

WARNING**Electrical Shock Hazard**

Death, serious injury, or electrical shock can result by contact with these circuits, devices and components. This instrument contains electrical circuits, devices and components operating at dangerous voltages. Always switch off the power to the system and disconnect the power cords before proceeding with any maintenance procedure that risks exposure to this hazard.

WARNING**Hot Surface Hazard**

Close contact with an operating or recently used infrared source can result in severe heat burns to the skin. Allow hot parts to cool before proceeding with any maintenance procedure.

WARNING**Moving Parts Hazard**

Danger to hands. To avoid exposing moving parts, always operate your peristaltic pump with the tubing installed and the pump head lowered. While the peristaltic pump is running, do not put your hands on, or near, the pump head.

WARNING**Laser Hazard**

This is a Class 2 laser product. Although not powerful enough to harm your skin should your hand intercept it, the laser light could cause retinal (eye) damage if prolonged direct viewing occurs. Always wear appropriate safety equipment and clothing.

NOTE

This section refers only to maintenance procedures for the Agilent Cary FTIR Oil Analyzer. You should refer to your computer and printer manuals for their maintenance procedures and to Resolutions Pro Help for the maintenance procedures for any other accessories you ordered for use on your spectrometer.

Routine

Routine maintenance information is provided in the Resolutions Pro Help.

To access the Maintenance section of the Help:

- 1 Start the Oil Analyzer software.
- 2 Click **Help** on the toolbar to display the Help.
- 3 Click the **Contents** tab if necessary.
- 4 Click the plus (+) icon next to 'Maintenance' to expand it.
- 5 Click FTIR Oil Analyzer.

Alternatively:

- 1** Start the Resolutions Pro software.
- 2** Press F1 on the keyboard to display the Help.
- 3** Follow Steps 3 to 5 above.

Cleaning

Any spills in the sample compartment or on the FTIR Oil Analyzer should be immediately wiped up.


The exterior surfaces of the instrument should be kept clean. All cleaning should be done with a soft cloth. If necessary, this cloth can be dampened with water or a mild detergent.

Do not use organic solvents or abrasive cleaning agents.

Spare parts

For spare parts and consumables ordering information, refer to the Agilent website, www.agilent.com

This page is intentionally left blank.



7. Troubleshooting

For troubleshooting advice, refer to the Agilent Resolutions Pro Help.

To access the Troubleshooting section of the Help:

- 1 Start the Oil Analyzer software.
- 2 Click **Help** on the toolbar to display the Help.
- 3 Click the **Contents** tab if necessary.
- 4 Click the 'plus' (+) icon next to 'Troubleshooting' to expand it.
- 5 Click **FTIR Oil Analyzer**.

Alternatively:

- 1 Start the Resolutions Pro software.
- 2 Press F1 on the keyboard to display the Help.

Follow Steps 3 to 5 above.

This page is intentionally left blank.